



## What's New Questionnaire Answers March 2016


**A great big THANK YOU to all of you who supplied feedback from your experiences at WNU 2016. Please visit our Facebook page to view and share photos and memories! We look forward to having you back as a part of What's New University in 2018!**

Congratulations to:

Nissa Franke - St. Cloud Opticians

Andrea Ruane - Eye Candy

Alysha Sperry - Eye & Vision Clinics - De Pere



1) **T or F? A** Binocular PD is NOT given as one number. **A binocular PD is given as ONE number.**

2) Multifocal height is determined by measuring:

- A. Lower Edge of the Lens**
- B. Lower Edge of the Frame
- C. Upper Edge of the Lens
- D. Upper Edge of the Frame

3) When fitting a frame, you want the tilt to be \_\_\_?

- A. Retrosopic
- B. Pontoscopic**
- C. Positive
- D. Orhoscopic

4) **T or F? T** The index of refraction of the cornea is 1.58. **The index of refraction of the cornea is 1.376.**

5) What is the minimum thickness for CR-39 (dress)?

- A. 7.5 mm
- B. 1.0 mm
- C. 2.0 mm**
- D. 2.2 mm

**This question caused a lot of good debate. The target is 2.0 mm, but ANSI states that the center thickness shall not deviate from the nomial value more than +/- .3mm. Therefore, though 2.2 is in that range, 2.0 is still the minimum approved listed.**

6) Recommendations/Updates for What's New University in 2018?